

CATALC	1 #	TYPE
JOB NA	E WATTAGE	VOLTAGE

6" 1 PAR30 75 WATT MAXIMUM RR00622

MRI Non-Ferrous 100 CRI Lighting with Attractive Alzak® Low Brightness Trim

Features

Non-Magnetic Lamp/Socket

- · Designed for tungsten halogen GE or Osram-Sylvania lamp: 75 watt (max.) PAR30S (by others). Rated
- 100 CRI.
- Dimmable (controls by others).

2000-2500 hour life.

- For operation on D.C. circuits.
- · Glazed porcelain medium base vertical socket. Nickel plated brass screw shell. Silicone leads.

- \bullet Self-flanged parabolic SofTex $^{\!\top\!\!M}$ softly etched Alzak® aluminum reflector. Resists smudges, fingerprints, and dust.
- · Optional self-flanged colors and finishes available. See Options.

Housing

- · Acrylic enameled 100% aluminum.
- Rustproof: Exceeds 1000 spray test.

- · Cool: Dissipates heat across entire surface area.
- Entire luminaire visible and serviceable from below through removable reflector.
- Built-in plaster frame. Also see Option -79. **Outlet Box**
- Prewired. 100% aluminum. UL listed. • Accessible. 1/2" and 3/4" knockouts.

Installation

- · Lightweight. Recesses in any ceiling type
- 27" aluminum hanger bars supplied (2
- Fully adjustable universal non-mag brackets supplied (2).

UL, C-UL (Canada) Listing

- dry locations, covered cei<mark>lings</mark>
- circuit condu

Year Limited Warran


Thermal Protection (Per Current NEC)

Thermal protector included which provide automatic cut-off if rated temperatur exceeded. (Insulation related)


Recessed Round Open MRI


at a Glance


RR00622



Ceiling Cut-out 61/2"









Dia. (in ft.) shown is where FC value is half the FC at nadir.

Cone of Light Key Distance from fixture Footcandles at nadir (0°)

Circle of light at 50% of FC

THE KIRLIN COMPANY

3401 EAST JEFFERSON AVENUE • DETROIT, MICHIGAN 48207-4232 (313) 259-6400 • Fax: (313) 259-9409 OR (313) 259-3121 • www.kirlinlighting.com

HEALTHCARE FACILITY & PATIENT LIGHTING

MAXIMUM WATTAGE

1 PAR30 75

CATALOG NUMBER

RR00622




LIMITED WARRANTY: CATALOGED KRILIN FIXTURES ARE WARRANTIED FREE OF DEFECTS IN WORKMANSHIP OR MATERIAL FOR THREE price on presention of proof by purbase and defeatwhenture at the offices of manufacture within these years of orginal shipment. Liability of manufacture is limited to the foregoing warranties, expressed or implied. Safer does not warrant as to merchantability of liness for a particular use, norwal may not assistent not extend the awarranty or amend this specific warranty.

YEARS FROM DATE OF PURCHASE, INSTALLED TO N.E.C., IN NORMAL USE, Man, and manufacturer shall not be liable for any other damages, other to indirect, sustained on suffered by purior rank, Abnormal use, abuse of the fature, impoper installation, or repair, or modification of the fature without

Detailed Photometry - Installed Fixture

Photometry from I.T.L., Boulder, CO

RR00622 (75 Watt Narrow Flood)



EFFICIENCY: 32.0%

	CANDLEPOWER DISTRIBUTION								
		0.0	22.5	45.0	67.5	90.0			
	0	4992	4992	4992	4992	4992			
	5	4083	4083	4083	4083	4083			
1	5	1727	1727	1727	1727	1727			
2	5	309	309	309	309	309			
3	5	99	99	99	99	99			
4	5	16	16	16	16	16			
5	5	1	1	1	1				
6	5	0	0	0_	0	0			
7	5	0	0	0	0	0			
8	5	0	0	0	0	. 0			
9	0	0	0	0	0	0			

COEFFICIENTS OF UTILIZATION								
RF	20							
RC	80				50]
RW	70	50	30	7	_ \	30	10]
1	114	112	110	108	106	104	103	
2	110	107	104	101	102	100	98	
2	107	102	98	95	98	95	93	
V	103	98	94	91	95	92	89	
5	100	94	90	87	92	88	86	
6	97	91	87	84	89	85	83	
7	94	88	84	81	86	83	80	
8	92	85	81	78	84	80	77	L
	89	82	78	75	81	78	75	
10	87	80	76	73	79	75	73	

TL TEST #46932

Options

- -30 All white self-flanged reflector.
- -31 White flange (only) on self-flanged reflector.
- -32 Trim extender. Broadens trim flange O.D. by 1" to cover larger ceiling opening.
- -45 Gasket above reflector flange. (Minimizes dust streaks on ceiling.)
- -62 Clear specular Alzak® self-flanged reflector.
- -79 Extension collar for up to 2" thick ceiling.
- -94 SofTex or specular custom color / finish self-
- flanged reflector. Consult factory.
- 99 Special modification. Consult factory.
- P With PAR lamp. Specify wattage, style and beam spread.

SUBMITTAL DATA

APPROVAL STAMP

JOB NAME

TYPE

WATTAGE VOLTAGE

CATALOG NUMBER